

Human PLXNB1 ELISA Kit

Sandwich ELISA

Catalog EL-213-16122401

INTRODUCTION

PLXNB1 encodes Plexin-B1, a transmembrane receptor for the semaphorin family ligand SEMA4D/CD100; ligand binding activates the small GTPase RHOA to remodel the actin cytoskeleton, altering cell shape and loosening cell-cell adhesion contacts. On neurons, the same signaling axis, engaged by SEMA4A as well as SEMA4D, contributes to the wiring of inhibitory GABA-releasing synapses and steers axons along their intended paths during development. Plexin-B1 rarely acts alone at the membrane; it partners with co-receptors such as the neuropilins, the MET tyrosine kinase, and ErbB-2 to assemble larger signaling complexes that influence blood vessel growth and the organization of epithelial tissue. Because these pathways can be co-opted during malignant transformation, altered Plexin-B1 signaling has also been associated with tumor growth and with abnormal cell migration and invasion.

PRINCIPLE OF THE ASSAY

This kit uses a sandwich enzyme immunoassay format to detect and quantify Human PLXNB1. The microtiter plate wells are pre-coated with a PLXNB1-specific capture antibody that binds the target analyte present in standards or samples. A biotin-conjugated detection antibody, also specific to Human PLXNB1, is then added and binds a second epitope on the captured analyte. Avidin-conjugated horseradish peroxidase (HRP) is subsequently introduced and binds the biotin, completing the detection complex. Upon addition of TMB substrate solution, a colorimetric signal develops exclusively in wells where PLXNB1, the biotin-conjugated detection antibody, and the avidin-HRP conjugate are all present. The enzymatic reaction is stopped by the addition of a sulphuric acid stop solution, and absorbance is measured spectrophotometrically at 450 nm (± 10 nm). Sample PLXNB1 concentrations are then calculated by referencing the sample optical density values against a standard curve.

MATERIALS AND STORAGE

Store kit components at -20°C unless specified otherwise. DO NOT USE past kit expiration date. Some vials contain a small amount of reagents. Spin tubes on pulse setting prior to opening.

Each kit includes:	Units
Pre-Coated Microplate	12 strips x 8 wells
Standard (Lyophilized)	2 vials
Biotinylated Antibody (100 $\times$)	120 μL
Streptavidin-HRP (100 $\times$)	120 μL
Standard/Sample Diluent Buffer	20 mL
Biotinylated Antibody Diluent	12 mL
HRP Diluent	12 mL
Wash Buffer (25 $\times$)	20 mL
TMB Substrate Solution	10 mL
Stop Reagent	6 mL
Plate Covers	2 pieces
<i>Do not mix or substitute reagents with those from other lots.</i>	

MATERIALS AND INSTRUMENTS REQUIRED BUT NOT SUPPLIED

- Precision pipettes calibrated to deliver 5-1000 μL
- Multi-channel pipette calibrated to deliver 50-200 μL
- Plate shaker
- Disposable tips
- Vortex-Mixer
- Distilled or de-ionized water
- Microplate reader capable of reading 450nm with background subtraction at 620nm

SAFETY PRECAUTIONS

- The test protocol must be followed strictly.
- All reagents containing human material should be handled as if potentially infectious. Operators should wear gloves and protective

clothing when handling any patient sera or serum based products.

- The kit reagents contain antimicrobial agents, acid and 3,3',5,5'-tetramethylbenzidine. Avoid contact with the skin and eyes. Rinse immediately with plenty of water if any contact occurs.
- Any liquid that has been brought into contact with potentially infectious material has to be discarded in a container with a disinfectant. Disposal must be performed in accordance with local regulations.
- Disposal must be performed in accordance with local regulations.
- Only trained laboratory personnel should execute this test.

Human PLXNB1 ELISA Kit

Sandwich ELISA

Catalog EL-213-16122401

PREPARATION OF REAGENTS

Prepare only the appropriate amount of required reagent on the day of use. Store all reagents as per instructions stated on the label.

- 1. Wash Buffer (1X) Preparation: Dilute the 25X wash buffer concentrate 1:25 with double-distilled Water.**
- 2. Detection Reagent (1X) Preparation: Dilute the biotinylated antibody and streptavidin-HRP concentrates 100-fold with Biotinylated Antibody Diluent and HRP Diluent, respectively, immediately before use.**
- 3. Preparation of Calibrators: Prepare calibrators by serial dilution from a 100 down to 1.57 (plus a 0 blank), per the source manual.**

Sol'n ID	Source	Source Vol (μL)	Cal. Diluent (μL)	Final Vol (μL)	Final Conc. (ng/mL)
Std 1	Stock Standard	500	500	1000	100
Std 2	Std 1	500	500	1000	50
Std 3	Std 2	500	500	1000	25
Std 4	Std 3	500	500	1000	12.5
Std 5	Std 4	500	500	1000	6.25
Std 6	Std 5	500	500	1000	3.13
Std 7	Std 6	500	500	1000	1.57
Blank	Std 7	500	500	1000	0

SPECIMEN STORAGE

This kit is compatible with EDTA-plasma, heparin-plasma and serum samples. Samples can be stored at or below -20°C for up to 1 year.

ASSAY PROCEDURE

1. After the kit is equilibrated at room temperature, add 100 μL of Standard Working Buffer (gradually diluted according to the instructions) or 100 μL of sample to each well, and incubate at 37°C for 80 minutes.
2. Discard the liquid in the plate, add 200 μL 1× Wash Buffer to each well, and wash the plate 3 times. After pat it dry against clean absorbent paper, add 100 μL Biotinylated Antibody Working Solution (1×) to each well, incubate at 37°C for 50 minutes.
3. Discard the liquid in the plate, add 200 μL 1× Wash Buffer to each well, and wash the plate 3 times. After pat it dry against clean absorbent paper, add 100 μL 1× Streptavidin-HRP Working Solution to each well, incubate at 37°C for 50 minutes.
4. Discard the liquid in the plate, add 200 μL 1× Wash Buffer to each well, and wash the plate 5 times. After pat it dry against clean absorbent paper, add 90 μL TMB Substrate Solution to each well, incubate at 37°C for 20 minutes in the dark.
5. Add 50 μL Stop Solution to each well, shake plate on a plate shaker for 1 minute to mix. Record the OD at 450 nm immediately, calculation of the results.

CALCULATIONS AND RESULTS

1. Construct a standard curve by plotting the absorbance obtained from each standard against concentration. Use a 4 or 5 parameter curve fit. Alternatively a log-log curve fit may be used.
2. The concentration of the unknowns can be read directly from this standard curve using the absorbance value for each sample.

3. Any sample undiluted or diluted still reading greater than the highest standard should be diluted appropriately and retested. If the samples have been diluted, the concentration determined from the standard curve must be multiplied by the dilution factor.

PERFORMANCE CHARACTERISTICS

Precision: Intra-assay CV < 8%. Inter-assay CV < 10%.


Detection Limit: 0.64 ng/mL

Recovery: Serum (n = 5): 88-99% (avg 93%); EDTA plasma (n = 5): 86-98% (avg 92%); Heparin plasma (n = 5): 82-98% (avg 87%)

Specificity: This assay has high sensitivity and excellent specificity for detection of Human PLXNB1. No significant cross-reactivity or interference between Human PLXNB1 and analogues was observed.

Human PLXNB1 ELISA Kit Sandwich ELISA Catalog EL-213-16122401

SAMPLE STANDARD CURVE



Representative standard curve shown for illustration only. This curve should not be used to calculate unknown sample concentrations -- run a new standard curve with each assay.

ORDERING INFORMATION

Please visit www.affinityimmuno.com to order this product. Visa, Mastercard, AMEX and PayPal are accepted in our online store.

Your order will be processed immediately and you will be notified with a delivery timeframe.