This kit employs a sandwich enzyme immunoassay technique to quantify TSLPR (CRLF2) in biological samples.

Human TSLPR ELISA Kit
Sandwich ELISA
Catalog EL-213-202

INTRODUCTION

CRLF2 encodes the receptor subunit for thymic stromal lymphopoietin (TSLP), commonly referred to as TSLPR. Together with the interleukin-7 receptor chain, this receptor pairs with TSLP to trigger STAT3, STAT5 and JAK2 signaling, driving proliferation of hematopoietic progenitors and shaping normal blood cell development. Chromosomal rearrangements that place CRLF2 next to the immunoglobulin heavy-chain locus or the P2RY8 gene are recurrently found in B-progenitor acute lymphoblastic leukemia, including cases arising in children with Down syndrome. Because of this link, measuring CRLF2/TSLPR protein levels is of interest in research on leukemogenesis and lymphoid progenitor biology.

PRINCIPLE OF THE ASSAY

This kit uses a sandwich enzyme immunoassay format in which the microtiter plate wells are pre-coated with a capture antibody specific to Human CRLF2 (TSLPR). Standards or samples are pipetted into the appropriate wells along with a biotin-conjugated detection antibody that also recognizes Human CRLF2. Avidin conjugated to horseradish peroxidase (HRP) is then added to each well and allowed to incubate, linking the enzymatic reporter to the captured analyte via the biotin–avidin interaction. Upon addition of TMB substrate solution, a colorimetric signal develops exclusively in wells containing Human CRLF2, the biotin-conjugated antibody, and the avidin-HRP conjugate. The enzymatic reaction is stopped by the addition of sulphuric acid solution, and absorbance is read spectrophotometrically at 450 nm (± 10 nm). Sample concentrations are determined by comparing the optical density of each sample against a standard curve.

MATERIALS AND STORAGE

Store kit components at -20°C unless specified otherwise. DO NOT USE past kit expiration date. Some vials contain a small amount of reagents. Spin tubes on pulse setting prior to opening.

Each kit includes:	Units
Pre-Coated Microplate	12 strips x 8 wells
Standard (Lyophilized)	2 vials
Biotinylated Antibody (100×)	120 µL
Streptavidin-HRP (100×)	120 µL
Standard/Sample Diluent Buffer	20 mL
Biotinylated Antibody Diluent	12 mL
HRP Diluent	12 mL
Wash Buffer (25×)	20 mL
TMB Substrate Solution	10 mL
Stop Reagent	6 mL
Plate Covers	2 pieces
Do not mix or substitute reagents with those from other lots.	

MATERIALS AND INSTRUMENTS REQUIRED BUT NOT SUPPLIED

- Precision pipettes calibrated to deliver 5-1000µL
- Multi-channel pipette calibrated to deliver 50-200µL
- Plate shaker
- Disposable tips
- Vortex-Mixer
- Distilled or de-ionized water
- Microplate reader capable of reading 450nm with background subtraction at 620nm

SAFETY PRECAUTIONS

- The test protocol must be followed strictly.
- All reagents containing human material should be handled as if potentially infectious. Operators should wear gloves and protective

clothing when handling any patient sera or serum based products.

- The kit reagents contain antimicrobial agents, acid and 3,3’,5,5’-tetramethylbenzidine. Avoid contact with the skin and eyes. Rinse immediately with plenty of water if any contact occurs.
- Any liquid that has been brought into contact with potentially infectious material has to be discarded in a container with a disinfectant. Disposal must be performed in accordance with local regulations.
- Disposal must be performed in accordance with local regulations.
- Only trained laboratory personnel should execute this test.

This kit employs a sandwich enzyme immunoassay technique to quantify TSLPR (CRLF2) in biological samples.

Human TSLPR ELISA Kit
Sandwich ELISA
Catalog EL-213-202

PREPARATION OF REAGENTS

Prepare only the appropriate amount of required reagent on the day of use. Store all reagents as per instructions stated on the label.

1. **Wash Buffer (1X) Preparation: Dilute the 25X wash buffer concentrate 1:25 with double-distilled Water.**
2. **Detection Reagent (1X) Preparation: Dilute the biotinylated antibody and streptavidin-HRP concentrates 100-fold with Biotinylated Antibody Diluent and HRP Diluent, respectively, immediately before use.**
3. **Preparation of Calibrators: Prepare calibrators by serial dilution from a 2000 down to 31.25 (plus a 0 blank), per the source manual.**

Sol'n ID	Source	Source Vol (µL)	Cal. Diluent (µL)	Final Vol (µL)	Final Conc. (ng/mL)
Std 1	Stock Standard	500	500	1000	2000
Std 2	Std 1	500	500	1000	1000
Std 3	Std 2	500	500	1000	500
Std 4	Std 3	500	500	1000	250
Std 5	Std 4	500	500	1000	125
Std 6	Std 5	500	500	1000	62.5
Std 7	Std 6	500	500	1000	31.25
Blank	Std 7	500	500	1000	0

SPECIMEN STORAGE

This kit is compatible with EDTA-plasma, heparin-plasma and serum samples. Samples can be stored at or below -20°C for up to 1 year.

ASSAY PROCEDURE

1. After the kit is equilibrated at room temperature, add 100 µL of Standard Working Buffer (gradually diluted according to the instructions) or 100 µL of sample to each well, and incubate at 37°C for 80 minutes.

2. Discard the liquid in the plate, add 200 µL 1× Wash Buffer to each well, and wash the plate 3 times. After pat it dry against clean absorbent paper, add 100 µL Biotinylated Antibody Working Solution (1×) to each well, incubate at 37°C for 50 minutes.

3. Discard the liquid in the plate, add 200 µL 1× Wash Buffer to each well, and wash the plate 3 times. After pat it dry against clean absorbent paper, add 100 µL 1× Streptavidin-HRP Working Solution to each well, incubate at 37°C for 50 minutes.

4. Discard the liquid in the plate, add 200 µL 1× Wash Buffer to each well, and wash the plate 5 times. After pat it dry against clean absorbent paper, add 90 µL TMB Substrate Solution to each well, incubate at 37°C for 20 minutes in the dark.

5. Add 50 µL Stop Solution to each well, shake plate on a plate shaker for 1 minute to mix. Record the OD at 450 nm immediately, calculation of the results.

CALCULATIONS AND RESULTS

1. Construct a standard curve by plotting the absorbance obtained from each standard against concentration. Use a 4 or 5 parameter curve fit. Alternatively a log-log curve fit may be used.
2. The concentration of the unknowns can be read directly from this standard curve using the absorbance value for each sample.

3. Any sample undiluted or diluted still reading greater than the highest standard should be diluted appropriately and retested. If the samples have been diluted, the concentration determined from the standard curve must be multiplied by the dilution factor.

PERFORMANCE CHARACTERISTICS

Precision: Intra-assay CV < 8%. Inter-assay CV < 10%.

Detection Limit: 11.3 pg/mL

Recovery: Serum (n = 5): 86-94% (avg 90%); EDTA plasma (n = 5): 93-106% (avg 100%); Heparin plasma (n = 5): 96-108% (avg 102%)

Specificity: This assay has high sensitivity and excellent specificity for detection of Human CRLF2. No significant cross-reactivity or interference between Human CRLF2 and analogues was observed.

This kit employs a sandwich enzyme immunoassay technique to quantify TSLPR (CRLF2) in biological samples.

Human TSLPR ELISA Kit
Sandwich ELISA
Catalog EL-213-202

SAMPLE STANDARD CURVE



Representative standard curve shown for illustration only. This curve should not be used to calculate unknown sample concentrations -- run a new standard curve with each assay.

ORDERING INFORMATION

Please visit *www.affinityimmuno.com* to order this product. Visa, Mastercard, AMEX and PayPal are accepted in our online store.

Your order will be processed immediately and you will be notified with a delivery timeframe.