

Human HMGB1 ELISA Kit

Sandwich ELISA

Catalog EL-213-813701

INTRODUCTION

HMGB1 is a nonhistone chromatin protein that, inside the nucleus, helps bend and organize DNA to support replication, transcription, chromatin remodeling, and repair processes such as V(D)J recombination. When cells are stressed, dying, or activated, HMGB1 can be released passively from necrotic or pyroptotic cells or actively secreted by immune cells, and outside the cell it behaves as a damage-associated molecular pattern that signals tissue injury to the immune system. Depending on its oxidation state, extracellular HMGB1 can act more like a chemokine that recruits inflammatory cells, a cytokine that triggers NF- κ B-driven release of factors such as TNF and IL-6 through receptors including RAGE and Toll-like receptors, or a tolerogenic signal associated with apoptotic clearance. Because of this dual nuclear and extracellular role, HMGB1 has been linked to processes ranging from normal genome maintenance to tumor cell migration and the broader inflammatory response seen in sepsis, trauma, and other acute injury states.

PRINCIPLE OF THE ASSAY

This kit uses a sandwich enzyme immunoassay format in which the microtiter plate wells are pre-coated with a capture antibody specific to Human HMGB1. During incubation, HMGB1 present in standards or samples is captured by this immobilized antibody, followed by addition of a biotin-conjugated detection antibody that recognizes a distinct epitope on the bound analyte. Avidin-conjugated horseradish peroxidase (HRP) is then added to each well, where it binds the biotin label. Upon addition of TMB substrate solution, a colorimetric signal develops exclusively in wells containing Human HMGB1, biotin-conjugated antibody, and Avidin-HRP conjugate. The enzymatic reaction is stopped by addition of a sulphuric acid solution, and absorbance is measured at 450 nm ($\pm$ 10 nm). Sample concentrations are determined by comparing the optical density of each sample against a standard curve.

MATERIALS AND STORAGE

Store kit components at -20°C unless specified otherwise. DO NOT USE past kit expiration date. Some vials contain a small amount of reagents. Spin tubes on pulse setting prior to opening.

Each kit includes:	Units
Pre-Coated Microplate	12 strips x 8 wells
Standard (Lyophilized)	2 vials
Biotinylated Antibody (100 $\times$)	120 μ L
Streptavidin-HRP (100 $\times$)	120 μ L
Standard/Sample Diluent Buffer	20 mL
Biotinylated Antibody Diluent	12 mL
HRP Diluent	12 mL
Wash Buffer (25 $\times$)	20 mL
TMB Substrate Solution	10 mL
Stop Reagent	6 mL
Plate Covers	2 pieces
<i>Do not mix or substitute reagents with those from other lots.</i>	

MATERIALS AND INSTRUMENTS REQUIRED BUT NOT SUPPLIED

- Precision pipettes calibrated to deliver 5-1000 μ L
- Multi-channel pipette calibrated to deliver 50-200 μ L
- Plate shaker
- Disposable tips
- Vortex-Mixer
- Distilled or de-ionized water
- Microplate reader capable of reading 450nm with background subtraction at 620nm

SAFETY PRECAUTIONS

- The test protocol must be followed strictly.
- All reagents containing human material should be handled as if potentially infectious. Operators should wear gloves and protective

clothing when handling any patient sera or serum based products.

- The kit reagents contain antimicrobial agents, acid and 3,3',5,5'-tetramethylbenzidine. Avoid contact with the skin and eyes. Rinse immediately with plenty of water if any contact occurs.
- Any liquid that has been brought into contact with potentially infectious material has to be discarded in a container with a disinfectant. Disposal must be performed in accordance with local regulations.
- Disposal must be performed in accordance with local regulations.
- Only trained laboratory personnel should execute this test.

Human HMGB1 ELISA Kit

Sandwich ELISA

Catalog EL-213-813701

PREPARATION OF REAGENTS

Prepare only the appropriate amount of required reagent on the day of use. Store all reagents as per instructions stated on the label.

- 1. Wash Buffer (1X) Preparation: Dilute the 25X wash buffer concentrate 1:25 with double-distilled Water.**
- 2. Detection Reagent (1X) Preparation: Dilute the biotinylated antibody and streptavidin-HRP concentrates 100-fold with Biotinylated Antibody Diluent and HRP Diluent, respectively, immediately before use.**
- 3. Preparation of Calibrators: Prepare calibrators by serial dilution from a 4000 down to 62.5 (plus a 0 blank), per the source manual.**

Sol'n ID	Source	Source Vol (μL)	Cal. Diluent (μL)	Final Vol (μL)	Final Conc. (ng/mL)
Std 1	Stock Standard	500	500	1000	4000
Std 2	Std 1	500	500	1000	2000
Std 3	Std 2	500	500	1000	1000
Std 4	Std 3	500	500	1000	500
Std 5	Std 4	500	500	1000	250
Std 6	Std 5	500	500	1000	125
Std 7	Std 6	500	500	1000	62.5
Blank	Std 7	500	500	1000	0

SPECIMEN STORAGE

This kit is compatible with EDTA-plasma, heparin-plasma and serum samples. Samples can be stored at or below -20°C for up to 1 year.

ASSAY PROCEDURE

1. After the kit is equilibrated at room temperature, add 100 μL of Standard Working Buffer (gradually diluted according to the instructions) or 100 μL of sample to each well, and incubate at 37°C for 80 minutes.
2. Discard the liquid in the plate, add 200 μL 1× Wash Buffer to each well, and wash the plate 3 times. After pat it dry against clean absorbent paper, add 100 μL Biotinylated Antibody Working Solution (1×) to each well, incubate at 37°C for 50 minutes.
3. Discard the liquid in the plate, add 200 μL 1× Wash Buffer to each well, and wash the plate 3 times. After pat it dry against clean absorbent paper, add 100 μL 1× Streptavidin-HRP Working Solution to each well, incubate at 37°C for 50 minutes.
4. Discard the liquid in the plate, add 200 μL 1× Wash Buffer to each well, and wash the plate 5 times. After pat it dry against clean absorbent paper, add 90 μL TMB Substrate Solution to each well, incubate at 37°C for 20 minutes in the dark.
5. Add 50 μL Stop Solution to each well, shake plate on a plate shaker for 1 minute to mix. Record the OD at 450 nm immediately, calculation of the results.

CALCULATIONS AND RESULTS

1. Construct a standard curve by plotting the absorbance obtained from each standard against concentration. Use a 4 or 5 parameter curve fit. Alternatively a log-log curve fit may be used.
2. The concentration of the unknowns can be read directly from this standard curve using the absorbance value for each sample.

3. Any sample undiluted or diluted still reading greater than the highest standard should be diluted appropriately and retested. If the samples have been diluted, the concentration determined from the standard curve must be multiplied by the dilution factor.

PERFORMANCE CHARACTERISTICS

Precision: Intra-assay CV < 8%. Inter-assay CV < 10%.

Detection Limit: 26.9 pg/mL

Recovery: Serum (n = 5): 80-97% (avg 88%); EDTA plasma (n = 5): 85-101% (avg 93%); Heparin plasma (n = 5): 83-97% (avg 90%)


Specificity: This assay has high sensitivity and excellent specificity for detection of Human HMGB-1. No significant cross-reactivity or interference between Human HMGB-1 and analogues was observed.



This kit employs a sandwich enzyme immunoassay technique to quantify HMGB1 in biological samples.

Human HMGB1 ELISA Kit Sandwich ELISA Catalog EL-213-813701

SAMPLE STANDARD CURVE



Representative standard curve shown for illustration only. This curve should not be used to calculate unknown sample concentrations -- run a new standard curve with each assay.

ORDERING INFORMATION

Please visit www.affinityimmuno.com to order this product. Visa, Mastercard, AMEX and PayPal are accepted in our online store.

Your order will be processed immediately and you will be notified with a delivery timeframe.